

Design of a Charge Measurement Device

DESIGN DOCUMENT

Sdmay20-11
Honeywell
Long Que

Scribe/Nicholas Wolf, Internal Meeting Facilitator/Daniel Frantik,
External Meeting Facilitator/Brandon Degelau, Test Engineer/Ben
Buettner, Chief Engineer/Keagan Plummer, Report Manager/Colin
Ishman

sdmay20-11@iastate.edu
<https://sdmay20-11.sd.ece.iastate.edu>

Revised: 12-4-19/Rev3

Executive Summary

Development Standards & Practices Used

IEEE standards

Summary of Requirements

Requirements:

- Can measure the charge between 10 nano-Coulombs and 300 nano-Coulombs
- Can operate in voltages between 250- and 750-volts DC
- Confirmation of our output's accuracy

Recommendations:

- Small scale prototype circuit by the end of first semester
- Final design size approximately 3" by 6" PCB

Applicable Courses from Iowa State University Curriculum

- EE 201
- EE 230
- EE 303
- EE 311
- EE 333
- ENGL 314

New Skills/Knowledge acquired that was not taught in courses

- High Voltage Safety
- Larger team professional interactions
- Working with a client
- Following a budget

Table of Contents

1	Introduction	4
1.1	Acknowledgement	4
1.2	Problem and Project Statement	4
1.3	Operational Environment	4
1.4	Requirements	4
1.5	Intended Users and Uses	5
1.6	Assumptions and Limitations	5
1.7	Expected End Product and Deliverables	5
2.	Specifications and Analysis	5
2.1	Proposed Design	5
2.2	Design Analysis	6
2.3	Development Process	6
2.4	Design Plan	6
3.	Statement of Work	6
3.1	Previous Work And Literature	7
3.2	Technology Considerations	7
3.3	Task Decomposition	7
3.4	Possible Risks And Risk Management	7
3.5	Project Proposed Milestones and Evaluation Criteria	7
3.6	Project Tracking Procedures	7
3.7	Expected Results and Validation	8
4.	Project Timeline, Estimated Resources, and Challenges	8
4.1	Project Timeline	8
4.2	Feasibility Assessment	9
4.3	Personnel Effort Requirements	9
4.4	Other Resource Requirements	10
4.5	Financial Requirements	10
5.	Testing and Implementation	11
5.1	Interface Specifications	11
5.2	Hardware and software	11
5.2	Functional Testing	11

5.3	Non-Functional Testing	11
5.4	Process	12
5.5	Results	12
6.	Closing Material	13
6.1	Conclusion	13
6.2	References	13
6.3	Appendices	13

List of figures/tables/symbols/definitions

Figure 1: Example Charge Measurement Circuitry

Figure 2: Timeline chart

Table 1: Timeline Descriptions

Figure 3: Charge Measurement Circuit.

1 Introduction

1.1 ACKNOWLEDGEMENT

The company we will be receiving both technical advice and financial aid from is Honeywell. Jacob Starr and Nathen Orton are the project design leads we will be in contact with throughout the year. Another source of technical advice is Long Que, our group's advisor through Iowa State University (ISU). Sources for equipment and other necessities have yet to be determined. We suspect that outside equipment may come from either the Electronics Technology Group (ETG), Honeywell, or Ames Laboratory.

1.2 PROBLEM AND PROJECT STATEMENT

Problem statement-

Honeywell needs a way to accurately measure the accumulation of charge on electronic devices they are testing. The desired device should be reasonably sized and capable of handling any conditions or operating points that Honeywell specifies. The application of this device is important for extending Honeywell's ability to accurately test and understand the behavior of the electronics they are developing.

Solution approach-

Our team is currently exploring potential solutions through both research of new information and recollection of relevant material from previous courses taken at ISU. We have created a process flow that will be referenced throughout the duration of the project. This will help everyone to stay organized and in sync with the current state of the design. An Agile-based process will also be of use to our team for organization and execution of all tasks.

1.3 OPERATIONAL ENVIRONMENT

The charge measurement device will be used in a lab/test environment. Since it will not be exposed to irregular or uncontrolled weather conditions, it is unnecessary to spend extensive time on an enclosure that will protect against such instances. However, in a lab setting there are still potential hazards such as dropping the device, unintentional use outside of its capabilities, and high voltage contact. The final product's enclosure will ideally be a solid, well fitted casing that will primarily protect users from high voltage contact while protecting the device from small daily wear and tear events. Overall the enclosure will remain simple, economic, and easily duplicatable.

1.4 REQUIREMENTS

- Capable of operating at 250-750 V
- Accurately measure charge accumulation from 10-300 nC
- Only DC operation
- Connect to a computer UI
- Easily duplicated/simple design
- Relatively small PCB (around 3"x6") and final product enclosure

- Enclosure should protect the user from the high voltage
- Remain within budget restrictions (\$2500 per semester)

1.5 INTENDED USERS AND USES

This design is purposed for company use; the user is expected to be experienced or knowledgeable regarding the functionality and application. The intended application for this device is to measure the accumulation of charge in a testing environment. Because of this, the information Honeywell can share with our team is limited. With a testing application, we anticipate that the device will be used multiple times on a regular basis, so the end-product should be durable, reliable, and safe.

1.6 ASSUMPTIONS AND LIMITATIONS

Assumptions-

- The end-product is only intended for company use at Honeywell
- Honeywell plans to duplicate the design and create more units
- Only intended for DC operation
- Device will only measure the charge on one pin at a time

Limitations-

- PCB should be small (around 3"x6")
- \$2500 budget for two year-long projects
- System must be capable of operating safely between 250-750V
- Prototype is expected at the end of first semester
- Accessibility to a high voltage power supply for testing

1.7 EXPECTED END PRODUCT AND DELIVERABLES

In the project proposal form, there were a few examples of expected deliverables listed. Since the product is not intended for commercialization, the deliverables expected are limited to a final presentation, final report, and any developed prototype(s). The final presentation will be an outline of the design process used throughout the course of the project. This includes a timeline of events, difficulties faced in the design and testing processes, methods of how the team overcame the obstacles, and a summary on the general success of the project. The presentation will near the end of the second semester. A final report will be included to document functionality of the final product in comparison to expectations, components used in the design, a list of costs accumulated through developing prototypes, etc. This will also be near the end of the second semester. The final product or prototype is expected to be an accurately functioning charge measurement device. A prototype will be developed by the end of the first semester. With the device, we will include a list of specifications that describe the basic functions and capabilities.


2. Specifications and Analysis

2.1 PROPOSED DESIGN

The design that we have decided to try to implement for our charge measurement circuit is simple. As seen in Figure 1, the circuit takes an input voltage compared to system ground. The input voltage

will be from the shell of the connector. This circuit will be able to tell us the charge on the shell by having a voltage across the capacitor. The Charge is equal to voltage multiplied by capacitance. This will then go into a buffer. The output of the buffer will be going into an ADC to be read digitally. The resistor is in place to limit the current going into the amplifier and the switch is to discharge any leftover charge between measurements. The charge will correspond directly with the value of C1, so we will choose this value depending on the accuracy the client requires.

Figure 1: Example Charge Measurement Circuitry



2.2 DESIGN ANALYSIS

The difficulty with this design come from working in a high voltage system. We will need to find components that are able to hold close to 1000 volts. Another difficulty that we will have will pertain to choosing our capacitor. The value of the capacitor needs to be as exact as possible to have the best accuracy in our measurement. We will also want to have a way to control the offset of the amplifier. This needs to be as precise as possible. Some pros of this circuit are that it is a simple design. This should give us the ability to replicate the circuit easily for all the pins in the system. We should also be able to find a plethora of parts to make the circuit as small as possible at an affordable price.

2.3 DEVELOPMENT PROCESS

We are following the Agile process for this project. We want to use this style because it encourages communication between the members of the group. We believe that to get the best product we all need to contribute and voice our opinions. Agile allows everyone to be a part of every step in the process and encourages discussion between everyone.

2.4 DESIGN PLAN

Our design plan for this project begins with researching different charge measurement methods. After weighing the pros and cons of the different circuits, we will begin testing smaller versions of the circuit. For us this will mean inputs with smaller voltages. We will transition this design to the higher voltages per the requirements. After running tests on this iteration, we will connect the circuit to SHV connectors. After testing this with one pin of the connector, we will attempt to expand to multiple pins of the connector if time permits. A PCB will be developed on this last iteration of the design.

3. Statement of Work

3.1 PREVIOUS WORK AND LITERATURE

Electrometers have been used since the 18th century but could only handle nanovolts. Modern day electrometers can handle much larger voltages. For DC, a vibrating reed electrometer is used to measure charge. These can handle much larger DC voltages and the output will supply a constant current to keep the capacitance constant. Value electrometers use vacuum tubes with a very high gain and input resistance. The value model has a very small leakage making it less accurate than most others designs.

3.2 TECHNOLOGY CONSIDERATIONS

The vibrating reed electrometer is the only electrometer that can used for DC voltage. The issues with this design are the size and cost of the capacitors. It also is not as accurate as other designs but those are for AC only.

3.3 TASK DECOMPOSITION

- Charge Measurement Research
- Low voltage testing
- Part research
- High voltage testing
- Pin expansion (optional)
- Computer UI
- PCB design
- Enclosure Design

3.4 POSSIBLE RISKS AND RISK MANAGEMENT

This project involves working with high DC voltages. This can be a dangerous to team members and to the components. The team will need to create safety procedures and redundancy to protect themselves and the project. This can include arc flash training, lock out tag out, and ground safety procedures.

3.5 PROJECT PROPOSED MILESTONES AND EVALUATION CRITERIA

The first milestone for this project will be the first circuit design. Once we have all agreed on a circuit, we will order parts for the low voltage model, this is our second milestone. The third milestone is testing our low voltage model until we are satisfied with it. For our fourth milestone we will order parts for our high voltage model. Testing this model will be our fifth milestone. Our last milestone will be finalizing our circuit.

3.6 PROJECT TRACKING PROCEDURES

Our team will be using Microsoft Teams to keep track of our progress. Within Microsoft Teams we will have a calendar that will have deadlines for tasks that we need to accomplish. There will also be a planner in Microsoft Teams that will allow us to keep track of tasks that we will accomplish soon.

3.7 EXPECTED RESULTS AND VALIDATION

A PCB that can measure the charge given the requirements of the customer. It should be easy and safe to use. Accuracy of the design will be determined while prototyping.

4. Project Timeline, Estimated Resources, and Challenges

4.1 PROJECT TIMELINE

Figure 2: Timeline chart

Project Step	Oct	Nov	Dec	Jan	Feb	Mar	Apr
Charge Measurement Research							
Low voltage model							
Part research							
PCB design							
High voltage model							
Pin expansion (optional)							
Computer UI							
Enclosure Design							

Blue – Research and Design

Green – Construction and Testing

Purple – Finalization

This timeline is being proposed as we do not currently have a lot of experience with the concept in general, searching for appropriate parts, and using high voltage. A lot of time will be spent researching the principles of charge measurement. We will use a low voltage model to safely develop testing and usage procedures. During this time, appropriate parts will be selected for the high voltage model. After understanding how charge measurement works, we will scale up the voltage to meet the requirements of the project. After a few months with that model and the accuracy and safety have been thoroughly verified, the team will break up into subgroups to finalize the product. A lot of time to finalize is desired so changes may still be made if necessary.

4.2 FEASIBILITY ASSESSMENT

The final product will be a finished PCB with optional enclosure. The PCB will, regardless of enclosure, have all high voltage circuitry protected for safety reasons. A jack will be mounted to the board for connecting to a computer. The PCB will be able to measure charge and store little if any charge. If the PCB does store charge, it shall be measurable.

4.3 PERSONNEL EFFORT REQUIREMENTS

Table 1: Timeline Descriptions

Task	Personnel Effort
Charge Measurement Research	Continuous Research on charge measurement until low-voltage model has been developed. All members are expected to understand the principles of how charge can be measured and weigh the advantages and disadvantages of different methods. This will be done on an individual basis with the preliminary design chosen as a group.
Low Voltage Model	As a team, the preliminary design will be tested on a breadboard prototype. This can be accomplished by just a few individuals, but testing will require the group's effort. Half of the group will focus primarily on this model and the whole group will work on testing.
Part Research	The other half of the group as described above will start investigating parts for the high voltage model. This group will give input for both models but will have more focus on selecting and ordering parts as necessary. This will be continuous until a working prototype for the high voltage model is developed.
PCB Design	Members of the team not working on part research will work on designing a PCB for the high voltage model. This will be a major factor at the beginning of the high voltage model and intermittently throughout that timeline.

High Voltage Model	After the low voltage model is developed and testing has shown good indications, work on the high voltage model will begin. There will still be ongoing research for parts as this task is being handled. Testing will also require the group's effort. This will be held back slightly by the PCB design group initially.
Pin Expansion (Optional)	A small amount of time will be dedicated to expanding to the other pins of the connector. Two personnel will be dedicated to this but will be less of a priority.
Computer UI	At this point, the design has entered the finalization stage. One person will be solely dedicated to this one aspect for the time specified in the Gantt chart.
Enclosure Design	This team will be formed after the PCB team is approaching their final design. Enclosure design members will work intermittently with the PCB design team.

4.4 OTHER RESOURCE REQUIREMENTS

- Breadboards
- Perforated Boards
- ICs necessary for low-voltage model
- Resistors necessary for low-voltage model
- High voltage rated ICs
- High voltage rated resistors
- Access to PCB design software
- Access to electrical simulation software
- Access to 3D modeling design software
- High voltage power supply
- Equipment capable of measuring high voltages
- Wires and connectors
- Soldering irons
- Solder
- Heat-shrink

4.5 FINANCIAL REQUIREMENTS

We are given \$2500 for each semester to work on this project. For the entire duration of the project, we will receive \$5000.

5. Testing and Implementation

5.1 Interface Specifications

The circuit will be tested on a small scale first semester using a breadboard. A multimeter can provide us the readings to verify that the concept behind the circuit is correct. In the second semester, once we have a working concept, the testing will move into the large-scale model. The GUI will need to be perfected so that testing can be safe and as hands-off as possible with these high voltages.

5.2 HARDWARE AND SOFTWARE

First Semester (small scale prototype testing)

- Multimeter
- DC power supply
- JUnit testing

Second Semester (large scale prototype testing)

- High voltage DC power supply
- GUI
- Multimeter

High voltage DC power supply – we will have to access this through Ames Laboratory or through Honeywell. Using this supply is mandatory for making sure our large-scale testing is functioning properly.

Multimeter – this will give us the voltage readings. This will help ensure that our concept is correct.

JUnit testing – This will test the computer GUI. This will help making sure that the correct value of the voltage is being displayed.

GUI – software that will enable the user to read the charge measurements. This will allow our users understand the measurements without having to interact directly with the high voltage circuit.

5.3 FUNCTIONAL TESTING

Each unit of our system will be tested individually prior to the integration of the circuit – the ADC, the GUI, etc.

5.4 NON-FUNCTIONAL TESTING

With non-functional testing, the quality of our system is the main concern. The user, Honeywell, should be able to apply voltages in the range specified with no undesired changes in the performance of our circuit.

Honeywell intends on using this circuit for multiple pins, so the testing extend to this once we have a functioning circuit that has been given many test cases.

5.5 PROCESS

There has not been any formal testing yet on both the low and high voltage models. However, the testing plan is devised as follows:


Low Voltage Model

This model will be very similar to the schematic shown below, although it will likely not include the gas discharge tube. All the components in this model will be through hole components and placed on a breadboard.

The voltage range for testing is restricted by the power supplies available to us in lab. Two power supplies will need to be used: One for powering the op amp and the ADC, the other for the coax cable, which will accumulate the charge on its conductive outer shield – this will connect to the reference capacitor.

The test voltage range will be 5 – 15 VDC. The multimeter will be used to confirm our voltages, and the charge will be calculated and compared to measurements. The switch will be used to discharge the capacitor after each test case. This will be reiterated for each increment.

Figure 3: Charge Measurement Circuit.



High Voltage Model

The testing for this model will follow a similar process as the low voltage model. Now, however, the voltage test range will be 250 – 750 VDC, so we will implement gas discharge tube to provide surge protection.

The main difference between this model's testing plan and the low voltage testing plan

is the use of a GUI. Ultimately, the user should not be interacting directly with this circuit, so the charge measurements need to be displayed for the user.

The software will undergo separate tests prior to being integrated into the rest of the system.

5.6 RESULTS

As of 12/2/19, we have not conducted any testing on either of the models. We plan to build our low voltage model before the end of this semester and test.

6. Closing Material

6.1 CONCLUSION

This semester we have been working with Honeywell for our senior design project. At the beginning of the semester we started researching our design. We also had problems getting our high voltage power supply at the beginning of the semester. After researching we designed a circuit. At the end of the semester we have been able to get our high voltage power supply and we are ordering parts for the low voltage model that we will be testing before we move to the high voltage model.

6.2 REFERENCES

This will likely be different than in the project plan, since these will be technical references versus the related work / market survey references. Do a professional citation style (ex. IEEE).